



Certificate of Analysis

Reference Gas Oil Sample No. 1


Agilent Part
Number: 5060-9086


Sample Lot
Number: CP-1111

ASTM D.02.04 Section H

(Proposed Revision of D2887-93 Fall 1996 Subcommittee 4 ballot)

% Off	°F	°C
IBP	239	115
5	304	151
10	349	176
15	393	201
20	435	224
25	469	243
30	499	259
35	526	275
40	552	289
45	576	302
50	594	312
55	610	321
60	629	332
65	649	343
70	669	354
75	690	365
80	712	378
85	736	391
90	764	407
95	803	428
FBP	887	475


Peter A. King, Ph.D.
VP, Technical Operations


Daniel J. Lamendola
Director of QA/RA