



## Certificate of Analysis

### Forensic Toxicology Comprehensive Mix – Submix 10B

**Agilent Part  
Number:** 5190-6167B

**Sample Lot  
Number:** CM-3206

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	Certified Value
Norbuprenorphine glucuronide	469887-29-4	RM06653	100.0 ± 0.5 µg/mL

**Solvent:** methanol

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

**Storage:** Store at <4°C, Do Not Freeze

**Date of release:** 11 June 2015

**Date of expiration:** 31 July 2016

  
William J. Leary  
Quality Assurance Manager