



Certificate of Analysis

HPLC Flushing Solvent

**Agilent Part
Number:** G1969-85026

**Sample Lot
Number:** CM-3691

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	Certified Value
isopropyl alcohol	000067-63-0	RM10297	50 % (v/v)
acetonitrile	000075-05-8	RM10520	25 % (v/v)
cyclohexane	000110-82-7	RM10521	15 % (v/v)
dichloromethane	000075-09-2	RM10119	10 % (v/v)

Solvent: none

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: 09 July 2015
Date of expiration: 31 August 2019



William J. Leary
Quality Assurance Manager