Agilent Technologies

Certificate of Conformance

Product: Syringe

Batch No.	M01-F0391
Part Number	5190-1483
Description	Syringe, 10.0 uL, FN, bevel tip

Accuracy:	±1%
Reproducibility:	±1%

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.

MN-0435-S_RevD 07/09