Flame Ionization Detector Sample

Catalog Number: 18710-60170

Lot Number: S0458
Page: 1

This analytical standard was gravimetrically prepared, and the analyte concentrations were verified using high resolution gas chromatography and/or high performance liquid chromatography. The solution was prepared at the nominal concentration stated on the box label. The true value for each analyte, determined gravimetrically, is listed below.

Component	% by Weight*
n-tetradecane	0.033
n-pentadecane	0.033
n-hexadecane	0.033

Solvent: hexane

* Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.



Agilent Technologies
Innovating the HP Way

Manufactured for Agilent by ULTRA Scientific, Inc., USA