Agilent Technologies

ULTRA SCIENTIFIC Analytical Solutions

Certificate of Analysis

Test Std, DB-624 (Cap/Mega), 13 Peaks

Agilent Part Number:	**200-0113**	**Sample Lot Number:**	**CC-0977A**

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001:2000 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
n-heptane	000142-82-5	67H1344	1.26 ± 0.01 mg/mL
n-nonane	000111-84-2	4230191	2.02 ± 0.01 mg/mL
n-octane	000111-65-9	9612212	1.40 ± 0.01 mg/mL
methanol	000067-56-1	45135	1.27 ± 0.01 mg/mL
1-propanol	000071-23-8	11581PC	1.26 ± 0.01 mg/mL
pyridine	000110-86-1	01028CB	1.40 ± 0.01 mg/mL
2-butanone	000078-93-3	CA-02161PO	1.25 ± 0.01 mg/mL
acetone	000067-64-1	45174	1.27 ± 0.01 mg/mL
1,2-dichloropropane	000078-87-5	DC120777	2.03 ± 0.01 mg/mL
tetrachloroethylene	000127-18-4	PS00344BR	3.26 ± 0.02 mg/mL
chlorobenzene	000108-90-7	03148HZ	2.00 ± 0.01 mg/mL

Matrix: methylene chloride (dichloromethane)

Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: 23 April, 2007
Date of expiration: 31 May, 2009

Edward Fitzgerald- PHD
Edward Fitzgerald
Senior Chemist
Ultra Scientific, Inc