



Certificate of Analysis

Test Std, DB-FFAP/OV-351 (Megabore)

**Agilent Part
Number:**

200-0132

**Sample Lot
Number:**

CA-0307A

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001:2000 registered quality system. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
n-octadecane	000593-45-3	9203162	0.251 ± 0.001 mg/mL
n-nonadecane	000629-92-5	KR11121DR	0.251 ± 0.001 mg/mL
1-octanol	000111-87-5	KQ04244KQ	0.251 ± 0.001 mg/mL
hexanoic acid	000142-62-1	CY00527BY	0.251 ± 0.001 mg/mL
isovaleric acid	000503-74-2	CU07812MS	0.250 ± 0.001 mg/mL
2-methylnaphthalene	000091-57-6	EP0040687	0.251 ± 0.001 mg/mL
ethyl dodecanoate	000106-33-2	L190B	0.251 ± 0.001 mg/mL
methyl undecanoate	001731-86-8	NT00201	0.251 ± 0.001 mg/mL

Matrix: hexane

Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: 9 November, 2006
Date of expiration: 31 December, 2008

Edward Fitzgerald- PHD
Edward Fitzgerald
Senior Chemist
Ultra Scientific, Inc