

Certificate of Analysis

ISO 17034

Organochlorine Pesticides Standard

Product Number: PPM-608C-1**Page:** 1 of 2**Lot Number:** 0006492558**Lot Issue Date:** 07-Oct-2019**Expiration Date:** 31-Oct-2020

This ISO 17034 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	RM00635	20.1 ± 0.1 µg/mL
alpha-BHC	000319-84-6	RM14886	20.1 ± 0.1 µg/mL
beta-BHC	000319-85-7	rm15392	20.0 ± 0.1 µg/mL
delta-BHC	000319-86-8	RM09156	20.1 ± 0.1 µg/mL
gamma-BHC	000058-89-9	RM02694	20.1 ± 0.1 µg/mL
4,4'-DDD	000072-54-8	RM00615	100.4 ± 0.5 µg/mL
4,4'-DDE	000072-55-9	RM02892	20.0 ± 0.1 µg/mL
4,4'-DDT	000050-29-3	RM00618	100.4 ± 0.5 µg/mL
dieldrin	000060-57-1	RM15334	20.0 ± 0.1 µg/mL
endosulfan I	000959-98-8	RM00834	20.1 ± 0.1 µg/mL
endosulfan II	033213-65-9	RM02904	100.4 ± 0.5 µg/mL
endosulfan sulfate	001031-07-8	RM15389	100.5 ± 0.5 µg/mL
endrin	000072-20-8	RM14459	100.2 ± 0.5 µg/mL
endrin aldehyde	007421-93-4	RM04131	20.1 ± 0.1 µg/mL
heptachlor	000076-44-8	RM03111	20.0 ± 0.1 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM15293	20.1 ± 0.1 µg/mL
methoxychlor	000072-43-5	RM14186	20.1 ± 0.1 µg/mL

Matrix: methanol (methyl alcohol)**Storage:** Store at Room Temperature (15° to 30°C).

Certificate of Analysis

ISO 17034

Organochlorine Pesticides Standard

Product Number: PPM-608C-1

Page: 2 of 2

Lot Number: 0006492558

Lot Issue Date: 07-Oct-2019

Expiration Date: 31-Oct-2020

Analyte

CAS#

Analyte Lot

True Value

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



Monica Bourgeois
QMS Representative