

Certificate of Analysis

ISO 17034

Supplemental Cannabis Pesticide Standard

Product Number: SCA-211-1**Page:** 1 of 1**Lot Number:** 0006506465**Lot Issue Date:** 02-Dec-2019**Expiration Date:** 31-Dec-2020

This ISO 17034 Reference Material (RM) was manufactured and verified in accordance with Agilent Technologies ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
captan	000133-06-2	RM02848	100.0 ± 0.5 µg/mL
coumaphos	000056-72-4	RM12738	100.3 ± 0.5 µg/mL
dimethomorph	110488-70-5	RM14353	100.3 ± 0.5 µg/mL
fenhexamid	126833-17-8	RM15391	100.2 ± 0.5 µg/mL
pentachloronitrobenzene	000082-68-8	RM02477	100.1 ± 0.5 µg/mL
mevinphos	007786-34-7	RM14344	100.1 ± 0.5 µg/mL
spinetoram	187166-40-1	RM15280	100.2 ± 0.5 µg/mL

Matrix: acetonitrile**Storage:** Store Frozen (-25° to -10°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS RepresentativeISO 17034 Cert No.
AR-1936Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management SystemISO 17025 Cert No.
AT-1937