

Certificate of Analysis

Product Name: Organochlorine Pesticides Standard

Product Number: US-127-1

Lot Issue Date: 17-Jan-2020

Lot Number: 0006515776

Expiration Date: 31-Jan-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system, and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
aldrin	000309-00-2	RM00635	2006 $\pm$ 10 μ g/mL
alpha-BHC	000319-84-6	RM13413	2006 $\pm$ 10 μ g/mL
beta-BHC	000319-85-7	RM06986	2009 $\pm$ 10 μ g/mL
delta-BHC	000319-86-8	RM09156	2005 $\pm$ 10 μ g/mL
gamma-BHC	000058-89-9	RM02694	2006 $\pm$ 10 μ g/mL
cis-chlordane	005103-71-9	RM07243	2004 $\pm$ 10 μ g/mL
trans-chlordane	005103-74-2	RM02726	2006 $\pm$ 10 μ g/mL
4,4'-DDD	000072-54-8	RM00615	2004 $\pm$ 10 μ g/mL
4,4'-DDE	000072-55-9	RM02892	2007 $\pm$ 10 μ g/mL
4,4'-DDT	000050-29-3	RM06888	2006 $\pm$ 10 μ g/mL
dieldrin	000060-57-1	RM03956	2006 $\pm$ 10 μ g/mL
endosulfan I	000959-98-8	RM00834	2007 $\pm$ 10 μ g/mL
endosulfan II	033213-65-9	RM07039	2003 $\pm$ 10 μ g/mL
endosulfan sulfate	001031-07-8	RM12608	2006 $\pm$ 10 μ g/mL
endrin	000072-20-8	RM00557	2005 $\pm$ 10 μ g/mL
endrin aldehyde	007421-93-4	RM12540	2004 $\pm$ 10 μ g/mL
endrin ketone	053494-70-5	RM12603	1992 $\pm$ 10 μ g/mL

Certificate of Analysis

Product Number: US-127-1

Lot Number: 0006515776

heptachlor	000076-44-8	RM03111	2005 ± 10 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM13469	2005 ± 10 µg/mL
methoxychlor	000072-43-5	RM12528	2007 ± 10 µg/mL

Matrix: hexane/toluene (1:1)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert No.
AR-1936

RM was produced in accordance with TUV USA Inc registered ISO
9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

[www.agilent.com/quality/
CSD-QA-015.1](http://www.agilent.com/quality/CSD-QA-015.1)



ISO 17025 Cert
No. AT-1937