

**Certificate of Analysis**

Description : Amino Acid Standard (0.1 nmol/μl) 10/pk

Part.-No. : 5061-3332

Production date : 02-Mar 2020

Lot-No. : BCCC9774

Expiration date : 02-Mar 2022

No.	Component	Molecular Weight	Final Concentration (mg/ml)
01	<b>L-Alanine</b>	89.10	0.00891
02	<b>L-Arginine</b>	174.20	0.01742
03	<b>L-Aspartic Acid</b>	133.11	0.01331
04	<b>L-Cystine</b>	240.30	0.02403
05	<b>L-Glutamic Acid</b>	147.13	0.01471
06	<b>Glycine</b>	75.07	0.00751
07	<b>L-Histidine Hydrochloride Monohydrate</b>	209.63	0.02096
08	<b>L-Isoleucine</b>	131.18	0.01312
09	<b>L-Leucine</b>	131.18	0.01312
10	<b>L-Lysine Hydrochloride</b>	182.65	0.01827
11	<b>L-Methionine</b>	149.21	0.01492
12	<b>L-Phenylalanine</b>	165.19	0.01652
13	<b>L-Proline</b>	115.13	0.01151
14	<b>L-Serine</b>	105.09	0.01051
15	<b>L-Threonine</b>	119.12	0.01191
16	<b>L-Tyrosine</b>	181.19	0.01812
17	<b>L-Valine</b>	117.15	0.01172

**Purity Determination:**

Amino acids by Titration, Loss on drying, Residue on ignition, IR-Spectroscopy, MicroSelect-Test, Trace Metal and Trace Anion Determination

Hydrochloric acid by Titration, MicroSelect-Test, Trace Metal and Trace Anion Determination

**Raw materials**

All raw materials used to prepare this amino acid standard are of the highest available purity (&gt; 99%) and are routinely analyzed according to the above-mentioned purity-determinations.

**Manufacturing**

We employ precise measuring techniques in manufacturing this amino acid standard. Mass is determined with electronic balances capable of weighing to 0.0001 g and calibrated by the Swiss Office of Weights and Measures. Volume is determined in dedicated high-purity borosilicate volumetric flasks capable of measuring a 2000 ml-volume with an accuracy of 0.3%.

**Packaging and Storage**

The final solution is handled under argon, filled into 1 ml amber ampoules under inert gas (argon) and sealed. 10 finished ampoules are packaged into a set and stored at 4 °C.

**Stability**

Every individual lot of the product is subjected to reanalysis and the experience allows to set the shelf life to two years, if the product is stored as received at 4 °C. The guaranteed stability is not applicable to ampoules stored after opening, even if resealed.

**Analytical Quality Control**

The scope of the analytical testing procedures covers identity, purity, homogeneity, accuracy, function test in amino acid analysis and stability of the finished product

Description	Lot Analysis	Specifications
Aspect	clear, colorless liquid	clear, colorless liquid
Density (20/4)	1.000	1.00 ± 0.01 g/ml
Indes of Refraction (20/D)	1.334	1.334 ± 0.005
Amino Acid Analysis :	corresponds	corresponds
- Identity	corresponds	corresponds
- Purity	corresponds	corresponds
- Concentration Accuracy	corresponds	± 2.5% relative to primary standard

Buchs, 02.03.2020

Sigma-Aldrich Production GmbH  
Industriestrasse 25  
CH-9470 Buchs / Switzerland



.....  
Dr. A. Schneider, Chemist Quality Assurance Sr  
Quality Assurance, Buchs / Switzerland