

**Certificate of Analysis**

Description : Amino Acid Standard (0.010 nmol/μl) 10/pk

Part.-No. : 5061-3334

Production date : 02-Mar 2020

Lot.-No. : BCCC9775

Expiration date : 02-Mar 2022

| No. | Component                                    | Molecular Weight | Final Concentration (mg/ml) |
|-----|--|------------------|-----------------------------|
| 01  | <b>L-Alanine</b>                             | 89.10            | 0.00089                     |
| 02  | <b>L-Arginine</b>                            | 174.20           | 0.00174                     |
| 03  | <b>L-Aspartic Acid</b>                       | 133.11           | 0.00133                     |
| 04  | <b>L-Cystine</b>                             | 240.30           | 0.00240                     |
| 05  | <b>L-Glutamic Acid</b>                       | 147.13           | 0.00147                     |
| 06  | <b>Glycine</b>                               | 75.07            | 0.00075                     |
| 07  | <b>L-Histidine Hydrochloride Monohydrate</b> | 209.63           | 0.00210                     |
| 08  | <b>L-Isoleucine</b>                          | 131.18           | 0.00131                     |
| 09  | <b>L-Leucine</b>                             | 131.18           | 0.00131                     |
| 10  | <b>L-Lysine Hydrochloride</b>                | 182.65           | 0.00183                     |
| 11  | <b>L-Methionine</b>                          | 149.21           | 0.00149                     |
| 12  | <b>L-Phenylalanine</b>                       | 165.19           | 0.00165                     |
| 13  | <b>L-Proline</b>                             | 115.13           | 0.00115                     |
| 14  | <b>L-Serine</b>                              | 105.09           | 0.00105                     |
| 15  | <b>L-Threonine</b>                           | 119.12           | 0.00119                     |
| 16  | <b>L-Tyrosine</b>                            | 181.19           | 0.00181                     |
| 17  | <b>L-Valine</b>                              | 117.15           | 0.00117                     |

**Purity Determination:**

Amino acids by Titration, Loss on drying, Residue on ignition, IR-Spectroscopy, MicroSelect-Test, Trace Metal and Trace Anion Determination

Hydrochloric acid by Titration, MicroSelect-Test, Trace Metal and Trace Anion Determination

**Raw materials**

All raw materials used to prepare this amino acid standard are of the highest available purity (&gt; 99%) and are routinely analyzed according to the above-mentioned purity-determinations.

**Manufacturing**

We employ precise measuring techniques in manufacturing this amino acid standard. Mass is determined with electronic balances capable of weighing to 0.0001 g and calibrated by the Swiss Office of Weights and Measures. Volume is determined in dedicated high-purity borosilicate volumetric flasks capable of measuring a 2000 ml-volume with an accuracy of 0.3%.

**Packaging And Storage**

The final solution is handled under argon, filled into 1 ml amber ampoules under inert gas (argon) and sealed. 10 finished ampoules are packaged into a set and stored at 4 °C.

**Stability**

Every individual lot of the product is subjected to reanalysis and the experience allows to set the shelf life to two years, if the product is stored as received at 4 °C. The guaranteed stability is not applicable to ampoules stored after opening, even if resealed.

**Analytical Quality Control**

The scope of the analytical testing procedures covers identity, purity, homogeneity, accuracy, function test in amino acid analysis and stability of the finished product

| Description                | Lot Analysis            | Specifications                      |
|----------------------------|-------------------------|-------------------------------------|
| Aspect                     | clear, colorless liquid | clear, colorless liquid             |
| Density (20/4)             | 1.000                   | 1.00 ± 0.01 g/ml                    |
| Index of Refraction (20/D) | 1.334                   | 1.334 ± 0.005                       |
| Amino Acid Analysis :      | corresponds             | corresponds                         |
| - Identity                 | corresponds             | corresponds                         |
| - Purity                   | corresponds             | corresponds                         |
| - Concentration Accuracy   | corresponds             | ± 2.5% relative to primary standard |

Buchs, 02.03.2020

Sigma-Aldrich Production GmbH  
Industriestrasse 25  
CH-9470 Buchs / Switzerland



.....  
Dr. A. Schneider, Chemist Quality Assurance Sr  
Quality Assurance, Buchs / Switzerland