

Certificate of Analysis

**ICP-MS Tuning Solution
High Concentration**

Catalog Number: CICM-696

Lot Number: 0006525309

Lot Issue Date: 03/16/2020


Expiration Date: 04/30/2022

This Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO 17034 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM	
lithium	10.0	±	0.1	µg/mL	ICP / ICP-OES	3129a
yttrium	10.0	±	0.1	µg/mL	ICP / ICP-OES	3167a
cerium	10.0	±	0.1	µg/mL	ICP / ICP-OES	3110
thallium	10.0	±	0.1	µg/mL	ICP / ICP-OES	3158
cobalt	10.0	±	0.1	µg/mL	ICP / ICP-OES	3113

Matrix: 2% nitric acid in low TOC water (< 50 ppb)

Agilent uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of these standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.



Monica Bourgeois
QMS Representative



ISO 17034 Cert No.
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management System



ISO 17025 Cert No.
AT-1937