

Certificate of Analysis

EPA Method 6020A
Interference Check Standard A

Catalog Number: ICM-635
Lot Number: 0006525334
Lot Issue Date: 03/16/2020
Expiration Date: 04/30/2022

This Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO 17034 / ISO 17025 accredited laboratory and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM
aluminum	1001	± 5	µg/mL	ICP/ ICP-OES	3101a
carbon	2003	± 10	µg/mL	gravimetric	ANSI/NCSL Z-540-1
calcium	3005	± 15	µg/mL	ICP/ ICP-OES	3109a
chloride	21243	± 106	µg/mL	IC	3182
iron	2501	± 13	µg/mL	ICP/ ICP-OES	3126a
potassium	1001	± 5	µg/mL	ICP/ ICP-OES	3141a
magnesium	1001	± 5	µg/mL	ICP/ ICP-OES	3131a
molybdenum	20.0	± 0.1	µg/mL	ICP/ ICP-OES	3134
sodium	2501	± 13	µg/mL	ICP/ ICP-OES	3152a
phosphorus	1001	± 5	µg/mL	ICP/ ICP-OES	3139a
sulfur	1001	± 5	µg/mL	ICP/ ICP-OES	3154
titanium	20.0	± 0.1	µg/mL	ICP/ ICP-OES	3162a

Matrix: 5% nitric acid with trace hydrofluoric acid in low TOC water (< 50 ppb)

Agilent uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of these standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.


Monica Bourgeois
 QMS Representative



ISO 17034 Cert No.
 AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026
 registered ISO 9001 Quality Management System



ISO 17025 Cert No.
 AT-1937