

Certificate of Analysis

Product Name: GC Pesticide Standard #6

Product Number: PSM-100-F

Lot Issue Date: 18-May-2020

Lot Number: 0006538647

Expiration Date: 30-Jun-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system, and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
acclonifen	074070-46-5	RM07382	100.2 $\pm$ 0.5 μ g/mL
alachlor	015972-60-8	RM14112	100.4 $\pm$ 0.5 μ g/mL
chlorthiophos	060238-56-4	RM16218	100.4 $\pm$ 0.5 μ g/mL
etofenprox	080844-07-1	RM09144	100.3 $\pm$ 0.5 μ g/mL
fenvalerate	051630-58-1	RM09132	100.2 $\pm$ 0.5 μ g/mL
flamprop-isopropyl	052756-22-6	RM13224	100.4 $\pm$ 0.5 μ g/mL
flonicamid	158062-67-0	RM15236-01	100.4 $\pm$ 0.5 μ g/mL
halfenprox	111872-58-3	RM12773	100.3 $\pm$ 0.5 μ g/mL
isoprothiolane	050512-35-1	RM14282	100.4 $\pm$ 0.5 μ g/mL
pendimethalin	040487-42-1	RM16288	100.1 $\pm$ 0.5 μ g/mL
pentachloroanisole	001825-21-4	RM02457	100.3 $\pm$ 0.5 μ g/mL
picolinafen	137641-05-5	RM13888	100.4 $\pm$ 0.5 μ g/mL
propachlor	001918-16-7	RM15846	100.3 $\pm$ 0.5 μ g/mL
propanil	000709-98-8	RM03261	100.2 $\pm$ 0.5 μ g/mL
terbufos sulfone	056070-16-7	RM15184	100.4 $\pm$ 0.5 μ g/mL
triallate	002303-17-5	RM03385	100.1 $\pm$ 0.5 μ g/mL

Matrix: acetonitrile

Certificate of Analysis

Product Number: PSM-100-F

Lot Number: 0006538647

Storage Conditions: Store Refrigerated (2° - 8°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert No.
AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/



ISO 17025 Cert
No. AT-1937