

# Certificate of Analysis

## ICP-MS Internal Std Mix

**Agilent Part Number: 5188-6525**
**Lot Number: 54-123CRY2**

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Bi	7440-69-9	100 µg/mL	101 µg/mL	3106*	Bi	99.99+	Lu	7439-94-3	100 µg/mL	99.9 µg/mL	3130a*	Lu <sub>2</sub> O <sub>3</sub>	99.99+
Ge	7440-56-4	100 µg/mL	99.9 µg/mL	3120a*	GeO <sub>2</sub>	99.99+	Rh	7440-16-6	100 µg/mL	99.3 µg/mL	3144*	Rh(NO <sub>3</sub> ) <sub>3</sub> -H <sub>2</sub> O	99.99+
In	7440-74-6	100 µg/mL	101 µg/mL	3124a*	In	99.99+	Sc	7440-20-2	100 µg/mL	101 µg/mL	3148a*	Sc <sub>2</sub> O <sub>3</sub>	99.99+
Li*	7439-93-2	100 µg/mL	99.1 µg/mL	3129a*	<sup>6</sup> Li <sub>2</sub> CO <sub>3</sub>	99.99+	Tb	7440-27-9	100 µg/mL	100 µg/mL	3157a*	Tb <sub>4</sub> O <sub>7</sub>	99.99+

\* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

**Purity grades:**

Starting Materials: Shown above

Matrix:

 10% HNO<sub>3</sub>: HNO<sub>3</sub> (CAS No. 7697-37-2) high purity grade

**Traceability:**

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 52-229CR, 54-124CR, 52-080CR, 51-015CR.

**Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:**


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	0.2	Co	<0.1	Hf	0.4	Nb	0.3	Ru	0.4	Ti	<0.01
Al	10	Cr	<0.5	Hg	<0.08	Nd	0.03	Sb	0.3	Tl	0.4
As	2	Cs	<0.01	Ho	0.01	Ni	0.8	Se	<2	Tm	0.05
Au	<0.1	Cu	<0.6	Ir	<10	P	<100	Si	200	U	<0.02
B	10	Dy	0.09	K	10	Pb	1	Sm	<0.03	V	0.07
Ba	1	Er	0.07	La	0.09	Pd	0.1	Sn	<0.5	W	<0.2
Be	<0.2	Eu	0.04	Mg	6	Pr	<0.02	Sr	6	Y	0.6
Ca	60	Fe	9	Mn	0.08	Pt	<0.1	Ta	<0.02	Yb	<0.6
Cd	<0.2	Ga	0.1	Mo	<0.1	Rb	0.08	Te	<0.3	Zn	<0.5
Ce	0.06	Gd	0.03	Na	5	Re	0.1	Th	0.4	Zr	5

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

**Date of release:** May 31, 2020

**Date of expiration:** November 30, 2021


  
 QC Coordinator