



Certificate of Analysis

ICP-MS Internal Std Mix

Agilent Part Number: 5188-6525

Lot Number: 54-124CRY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Bi	7440-69-9	100 µg/mL	99.8 µg/mL	3106*	Bi	99.99+	Lu	7439-94-3	100 µg/mL	99.6 µg/mL	3130a*	Lu ₂ O ₃	99.99+
Ge	7440-56-4	100 µg/mL	99.5 µg/mL	3120a*	GeO ₂	99.99+	Rh	7440-16-6	100 µg/mL	99.2 µg/mL	3144*	Rh(NO ₃) ₃ ·H ₂ O	99.99+
In	7440-74-6	100 µg/mL	100 µg/mL	3124a*	In	99.99+	Sc	7440-20-2	100 µg/mL	99.4 µg/mL	3148a*	Sc ₂ O ₃	99.99+
Li ^e	7439-93-2	100 µg/mL	100 µg/mL	3129a*	⁶ Li ₂ CO ₃	99.99+	Tb	7440-27-9	100 µg/mL	99.9 µg/mL	3157a*	Tb ₄ O ₇	99.99+

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

10% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade**Traceability:**

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 51-015CR, 52-229CR, 54-123CR, 52-080CR.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	0.2	Co	<0.1	Hf	0.4	Nb	0.2	Ru	0.3	Ti	2
Al	5	Cr	0.8	Hg	<0.1	Nd	0.08	Sb	0.3	Tl	0.4
As	<6	Cs	0.07	Ho	0.03	Ni	0.8	Se	<7	Tm	0.04
Au	<0.1	Cu	1	Ir	<10	P	<100	Si	300	U	<0.02
B	10	Dy	0.05	K	6	Pb	1	Sm	<0.03	V	0.06
Ba	1	Er	0.06	La	0.1	Pd	0.1	Sn	<0.6	W	<0.2
Be	<0.2	Eu	0.05	Mg	5	Pr	<0.02	Sr	6	Y	0.6
Ca	40	Fe	9	Mn	<0.8	Pt	0.1	Ta	<0.02	Yb	<0.5
Cd	0.07	Ga	0.08	Mo	<0.1	Rb	0.08	Te	0.03	Zn	<0.2
Ce	0.08	Gd	0.06	Na	8	Re	0.08	Th	0.4	Zr	3

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

Date of release: August 15, 2020

Date of expiration: February 28, 2022

Brijender S. Tonk
 QC Coordinator