

Certificate of Analysis

Product Name: LC/MS Pesticide Standard #6

Product Number: 5190-0551-6

Lot Issue Date: 06-Oct-2020

Lot Number: 0006561826

Expiration Date: 30-Nov-2021

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
carbaryl	000063-25-2	RM16704	100.3 ± 0.5 µg/mL
trichlorfon	000052-68-6	RM15210	100.1 ± 0.5 µg/mL
propoxur	000114-26-1	RM16246	100.3 ± 0.5 µg/mL
trietazine	001912-26-1	RM03941	100.3 ± 0.5 µg/mL
2,3,5-trimethacarb	002655-15-4	RM15235	100.1 ± 0.5 µg/mL
carbosulfan	055285-14-8	RM16090	100.2 ± 0.5 µg/mL
diethofencarb	087130-20-9	RM09135	100.4 ± 0.5 µg/mL
fenoxycarb	072490-01-8	RM14810	100.2 ± 0.5 µg/mL
furathiocarb	065907-30-4	RM13653	100.4 ± 0.5 µg/mL
pencycuron	066063-05-6	RM17317	100.1 ± 0.5 µg/mL
pyrimethanil	053112-28-0	RM17169	100.1 ± 0.5 µg/mL
triadimenol	055219-65-3	RM15410	100.1 ± 0.5 µg/mL
flonicamid	158062-67-0	RM16744	100.1 ± 0.5 µg/mL
flumioxazin	103361-09-7	RM10516	100.4 ± 0.5 µg/mL
spiromesifen	283594-90-1	RM15237	100.3 ± 0.5 µg/mL
spirotetramat	203313-25-1	RM16444	100.2 ± 0.5 µg/mL
sulfentrazone	122836-35-5	RM16905	100.2 ± 0.5 µg/mL

Certificate of Analysis

Product Number: 5190-0551-6**Lot Number:** 0006561826

zoxamide	156052-68-5	RM16909	100.2 ± 0.5 µg/mL
omethoate	001113-02-6	RM14837	100.3 ± 0.5 µg/mL
mesosulfuron-methyl	208465-21-8	RM03233	100.1 ± 0.5 µg/mL
phosmet	000732-11-6	RM03255	100.2 ± 0.5 µg/mL

Matrix: acetonitrile**Storage Conditions:** Store Frozen (-25° to -10°C).**Traceability:**

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:
Monica Bourgeois
QMS RepresentativeISO 17034 Cert
No. AR-1936RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality
Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1ISO 17025 Cert
No. AT-1937