



Agilent Technologies

Certificate of Analysis

Pesticide Solution #8

Agilent Part

Number: AGI-T04-8

Sample Lot

Number: 0006518965

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte
异菌脲

Analyte
iprodione

CAS#
036734-19-7

Analyte Lot
RM14917

Certified Value
1003 ± 5 µg/mL

Solvent: acetonitrile

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001.

Storage: Store Refrigerated (2° to 8°C)

Date of release: 26 February 2020

Date of expiration: 30 September 2021

Monica Bourgeois
QMS Representative