



Agilent Technologies

Certificate of Analysis

Pesticide Solution #9

Agilent Part

Number:

AGI-T04-9

Sample Lot

Number:

0006518966

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte
毒杀芬

Analyte
toxaphene

CAS#
008001-35-2

Analyte Lot
RM15098

Certified Value
1002 ± 5 µg/mL

Solvent: acetonitrile

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Storage: Store Refrigerated (2° to 8°C)

Date of release: 26 February 2020

Date of expiration: 30 September 2021


Monica Bourgeois
QMS Representative