



Certificate of Analysis

Pesticide Test Mix Kit

**Agilent Part
Number:** G3440-85045

**Kit Lot
Number:** 0006557077

This analytical reference material kit was manufactured and verified in accordance with an ISO 9001 registered quality system.

Description	Part Number	Lot #
Pesticide Test Mix – Standard 1	AGI-P03-1	0006553492
Pesticide Test Mix – Standard 2	AGI-P03-2	0006553496
Pesticide Test Mix – Standard 3	AGI-P03-3	0006553498
Pesticide Test Mix – Standard 4	AGI-P03-4	0006553499
Pesticide Test Mix – Standard 5	AGI-P03-5	0006553491
Pesticide Test Mix – Standard 6	AGI-P03-6	0006553490
Pesticide Test Mix – Standard 7	AGI-P03-7	0006553500
Pesticide Test Mix – Standard 8	AGI-P03-8	0006553501
Pesticide Test Mix – Standard 9	AGI-P03-9	0006553503
Pesticide Test Mix – Standard 10	AGI-P03-10	0006553495

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Storage: Store Refrigerated (2° to 8° C)

Date of release: 27 August 2020

Date of expiration: 31 March 2021

To view the CofA for individual Test Mix Standards, please visit the Agilent Technologies website at <https://www.Agilent.com/en-us/library/CertificateofAnalysis>.


Monica Bourgeois
QMS Representative