

Certificate of Analysis

ICP-MS Tuning Solution 10 ppm 100mL

Agilent Part Number: 5190-0465
Lot Number: 12-223ABY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.1 µg/mL	3110*	Ce ₂ (CO ₃) ₃ ·x(H ₂ O)	99.99+	Mg	7439-95-4	10.0 µg/mL	9.98 µg/mL	3131a*	Mg	99.99+
Co	7440-48-4	10.0 µg/mL	9.88 µg/mL	3113*	Co	99.99+	Tl	7440-28-0	10.0 µg/mL	9.85 µg/mL	3158*	TlNO ₃	99.99+
Li	7439-93-2	10.0 µg/mL	10.0 µg/mL	3129a*	Li ₂ CO ₃	99.99+	Y	7440-65-5	10.0 µg/mL	10.0 µg/mL	3167a*	Y(NO ₃) ₃ ·6H ₂ O	99.99+

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

 2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 1-201YJ, 3-117MJ, 4-127MKB.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.01	Cs	<0.01	Ho	0.03	Nd	<0.01	Ru	<0.04	Te	<0.01
Al	<0.01	Cu	<0.04	In	<0.01	Ni	<0.07	Sb	<0.01	Th	0.01
As	<0.2	Dy	0.1	Ir	<0.01	P	<100	Sc	<0.1	Ti	<0.1
Au	<0.2	Er	<0.01	K	<10	Pb	<0.01	Se	<0.2	Tm	<0.01
B	0.4	Eu	<0.01	La	0.07	Pd	0.2	Si	<100	U	<0.01
Ba	0.1	Fe	1	Lu	<0.01	Pr	0.2	Sm	<0.01	V	<0.09
Be	<0.03	Ga	<0.1	Mn	0.4	Pt	<0.06	Sn	<0.01	W	<0.01
Bi	<0.01	Gd	<0.1	Mo	<0.01	Rb	<0.07	Sr	<0.07	Yb	<0.01
Ca	<0.2	Ge	<0.01	Na	<5	Re	0.01	Ta	<0.01	Zn	5
Cd	<0.01	Hf	<0.01	Nb	<0.01	Rh	0.02	Tb	<0.1	Zr	<0.06
Cr	<0.09	Hg	<0.2								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

Date of release: October 31, 2020

Date of expiration: April 30, 2022



 QC Coordinator