

Certificate of Analysis

ICP-MS Internal Std Mix

Agilent Part Number: 5188-6525
Lot Number: 55-089CRY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Bi	7440-69-9	100 µg/mL	99.5 µg/mL	3106*	Bi	99.99+	Lu	7439-94-3	100 µg/mL	98.7 µg/mL	3130a*	Lu ₂ O ₃	99.99+
Ge	7440-56-4	100 µg/mL	100 µg/mL	3120a*	GeO ₂	99.99+	Rh	7440-16-6	100 µg/mL	100 µg/mL	3144*	Rh(NO ₃) ₃ ·H ₂ O	99.99+
In	7440-74-6	100 µg/mL	99.0 µg/mL	3124a*	In	99.99+	Sc	7440-20-2	100 µg/mL	99.0 µg/mL	3148a*	Sc ₂ O ₃	99.99+
Li ^e	7439-93-2	100 µg/mL	99.4 µg/mL	3129a*	⁶ Li ₂ CO ₃	99.99+	Tb	7440-27-9	100 µg/mL	98.7 µg/mL	3157a*	Tb ₂ O ₃	99.99+

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

 10% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 54-080CR, 54-255CR, 54-123CR, 55-058CR.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.4	Co	<0.8	Hf	<0.3	Nb	0.1	Ru	<0.6	Ti	<2
Al	6	Cr	0.3	Hg	<0.6	Nd	<0.2	Sb	<0.3	Tl	<0.6
As	10	Cs	<0.1	Ho	0.1	Ni	<0.6	Se	<5	Tm	<0.1
Au	<0.7	Cu	<2	Ir	10	P	<500	Si	200	U	<0.08
B	6	Dy	<0.2	K	3	Pb	<2	Sm	<0.1	V	<0.6
Ba	0.9	Er	<0.2	La	0.06	Pd	<0.5	Sn	<1	W	<0.8
Be	<2	Eu	<0.05	Mg	3	Pr	<0.08	Sr	<4	Y	1
Ca	90	Fe	8	Mn	1	Pt	<0.2	Ta	<0.3	Yb	<0.7
Cd	<0.3	Ga	<0.9	Mo	<0.7	Rb	<1	Te	<4	Zn	0.9
Ce	0.09	Gd	<0.2	Na	4	Re	<0.05	Th	0.5	Zr	3

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

Date of release: November 30, 2020

Date of expiration: May 31, 2022



 QC Coordinator