

Certificate of Analysis

Product Name: Organochlorine Pesticide Standard

Product Number: PPM-130-1

Lot Issue Date: 05-Mar-2021

Lot Number: 0006592838

Expiration Date: 31-Jul-2021

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
alachlor	015972-60-8	RM14112	100.4 ± 0.5 µg/mL
aldrin	000309-00-2	RM00635	100.3 ± 0.5 µg/mL
atrazine	001912-24-9	RM17497	100.1 ± 0.5 µg/mL
lindane	000058-89-9	RM02694	100.3 ± 0.5 µg/mL
dieldrin	000060-57-1	RM17652	100.1 ± 0.5 µg/mL
endrin	000072-20-8	RM14459	100.2 ± 0.5 µg/mL
heptachlor	000076-44-8	RM03111	100.4 ± 0.5 µg/mL
hexachlorobenzene	000118-74-1	RM03125	100.4 ± 0.5 µg/mL
hexachlorocyclopentadiene	000077-47-4	RM12238	100.3 ± 0.5 µg/mL
methoxychlor	000072-43-5	RM14186	100.4 ± 0.5 µg/mL
cis-nonachlor	005103-73-1	RM03193	100.4 ± 0.5 µg/mL
trans-nonachlor	039765-80-5	RM03191	100.3 ± 0.5 µg/mL
cis-chlordane	005103-71-9	RM07243	100.4 ± 0.5 µg/mL
simazine	000122-34-9	RM00622	100.2 ± 0.5 µg/mL
trans-chlordane	005103-74-2	RM02726	100.3 ± 0.5 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM17614	100.4 ± 0.5 µg/mL

Matrix: methanol (methyl alcohol)

Certificate of Analysis

Product Number: PPM-130-1

Lot Number: 0006592838

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the LRQA registered ISO 9001:2015 Quality Management System. Cert # 10303760

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937