Certificate of Analysis

EPA Method 200.7

Plasma Solution

Catalog Number: ICM-237

Lot Number: 0006545425

Lot Issue Date: 06/24/2020

Expiration Date: 06/30/2022

This Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO 17034 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
arsenic	10.0	±	0.1	µg/mL	ICP / ICP-OES	3103a
lead	10.0	±	0.1	µg/mL	ICP / ICP-OES	3128
selenium	10.0	±	0.1	µg/mL	ICP / ICP-OES	3149
thallium	10.0	±	0.1	µg/mL	ICP / ICP-OES	3158

Matrix: 2% nitric acid in low TOC water (< 50 ppb)

Agilent uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of these standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Monica Bourgeois
QMS Representative

ANAB ACCREDITED ISO 17034 REFERENCE MATERIAL PRODUCER

ISO 17034 Cert No. AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management System

ANAB ACCREDITED ISO/IEC 17025 TESTING LABORATORY

ISO 17025 Cert No. AT-1937