

Certificate of Analysis

EPA Method 200.7 SIC Solution III

Catalog Number: ICM-243
Lot Number: CS-0624
Lot Issue Date: 02/12/2018
Expiration Date: 03/31/2022


This Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within $\pm 5.0\%$, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value	Analytical Method	NIST SRM
aluminum	30.1 $\pm$ 0.2 $\mu\text{g/mL}$	ICP / ICP-MS	3101a
iron	150.6 $\pm$ 0.8 $\mu\text{g/mL}$	ICP / ICP-MS	3126a
nickel	20.0 $\pm$ 0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3136

Matrix: 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.


 John Russo
 President


 Monica Bourgeois
 Director of QA/RA



ISO 9001 Registered Quality System – TUV USA