Certificate of Analysis

CLP ICP Interference Check Standard #1

Catalog Number: ICM-441
Lot Number: 0006542119
Lot Issue Date: 06/16/2020
Expiration Date: 07/31/2023

This Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO 17034 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
aluminum	5002	±	25	µg/mL	ICP / ICP-OES	3101a
calcium	5003	±	25	µg/mL	ICP / ICP-OES	3109a
iron	2003	±	10	µg/mL	ICP / ICP-OES	3126a
magnesium	5006	±	25	µg/mL	ICP / ICP-OES	3131a

Matrix: 5% nitric acid in low TOC water (< 50 ppb)

Agilent uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of these standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Monica Bourgeois
QMS Representative

ANAB ACCREDITED ISO 17034 REFERENCE MATERIAL PRODUCER
ISO 17034 Cert No. AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management System

ANAB ACCREDITED ISO/IEC 17025 TESTING LABORATORY
ISO 17025 Cert No. AT-1937