Certificate of Analysis

Product Name: Magnesium Standard

Product Number: ICP-012 **Lot Issue Date:** 22-Dec-2020

Lot Number: 0006578805 **Expiration Date:** 31-Jan-2028

Description:
This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system. The analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	Starting Material	Lot Number	Purity (%)	Analyte Concentration	Traceability and Method
magnesium	magnesium nitrate hexahydrate	RM13723	99.999	1000 ± 2 µg/mL	NIST SRM 3131a; ICP/OES

Solvent: water with dilute nitric acid

Non-Certified Values:

Density: 1.0132 g/mL @ 20.00 ± 0.05°C

Trace Metallic Impurities in Solution Standard in µg/mL:

*	Al	<0.005	ND	*	Ga	<0.005	ND	n	Nb			n	S		
*	Sb	<0.005	ND	n	Ge			n	Os			n	Ta		
*	As	<0.005	ND	n	Au			*	Pd	<0.005	ND	n	Te		
*	Ba	<0.005	ND	n	Hf			*	P	<0.005	ND	n	Tb		
*	Be	<0.005	ND	n	Ho			*	Pt	<0.005	ND	*	Tl	<0.005	ND
*	Bi	<0.005	ND	*	In	<0.005	ND	*	K	<0.005	ND	n	Th		
*	B	<0.005	ND	n	Ir			n	Pr			n	Tm		
*	Cd	<0.005	ND	*	Fe	<0.022	D	n	Re			*	Sn	<0.005	ND
n	Cs			*	La	<0.005	ND	n	Rh			*	Ti	<0.005	ND
*	Ca	<0.043	D	*	Pb	<0.005	ND	n	Rb			n	W		
n	Ce			*	Li	<0.005	ND	n	Ru			n	U		
*	Cr	<0.005	ND	n	Lu			n	Sm			*	V	<0.005	ND
*	Co	<0.005	ND	s	Mg			n	Sc			n	Yb		
*	Cu	<0.005	ND	*	Mn	<0.005	ND	*	Se	<0.005	ND	n	Y		
n	Dy			*	Hg	<0.005	ND	*	Si	<0.005	ND	*	Zn	<0.005	ND
*	Er	<0.005	ND	*	Mo	<0.005	ND	*	Ag	<0.005	ND	n	Zr		
*	Eu	<0.005	ND	n	Nd			*	Na	<0.005	ND				
*	Gd	<0.005	ND	*	Ni	<0.005	ND	*	Sr	<0.005	ND				

* - element checked for n - not checked for s - standard element
D - detected ND - not detected i - spectral interference

ISO 17034 Cert No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

www.agilent.com/quality/
CSD-QA-039.0

ISO 17025 Cert No. AT-1937

Certificate of Analysis

Product Number:	ICP-012	**Lot Issue Date:**	22-Dec-2020
Lot Number:	0006578805	**Expiration Date:**	31-Jan-2028

Storage: Store at Room Temperature (15° to 30°C).

Traceability:
Traceability has been established through an unbroken chain of comparisons, each having stated uncertainties. Comparisons are based on appropriate physical or chemical measurements, including gravimetric or volumetric dilution, where the mass or volume of a solution before and after dilution is measured. The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 819.

Homogeneity:
This RM was formulated and unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:
This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.□

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening and should be processed without delay for the true value to be valid within the stated uncertainties. Do not pipet from the bottle. Do not return any material removed for pipetting to the bottle. Tightly cap the bottle after removing any material and store according to the instructions noted above.

Hazards:
Refer to the Safety Data Sheet for information regarding this RM.

Expiration of Certification:
The certification of this RM is valid, within the measurement uncertainty specified, until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:
The certification of this RM is valid, within the measurement uncertainty specified, until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Sample lot approver:

Monica Bourgeois
QMS Representative

ANAB ACCREDITED ISO 17034 REFERENCE MATERIAL PRODUCER
ISO 17034 Cert No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

www.agilent.com/quality/
CSD-QA-039.0

ANAB ACCREDITED ISO/IEC 17025 TESTING LABORATORY
ISO 17025 Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 **www.agilent.com/quality**