

Certificate of Analysis

Product Name: GC Pesticide Standard #2

Product Number: PSM-100-B

Lot Issue Date: 15-Oct-2020

Lot Number: 0006565862

Expiration Date: 30-Nov-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
diazinon	000333-41-5	RM09202	100.0 $\pm$ 0.5 μ g/mL
dichlofenthion	000097-17-6	RM00208	100.3 $\pm$ 0.5 μ g/mL
dicrotophos	000141-66-2	RM16722	100.4 $\pm$ 0.5 μ g/mL
ditalimfos	005131-24-8	RM13865	100.3 $\pm$ 0.5 μ g/mL
edifenphos	017109-49-8	RM17087	100.1 $\pm$ 0.5 μ g/mL
ethion	000563-12-2	RM16971	100.3 $\pm$ 0.5 μ g/mL
etrimfos	038260-54-7	RM05539	100.0 $\pm$ 0.5 μ g/mL
famphur	000052-85-7	RM15347	100.4 $\pm$ 0.5 μ g/mL
fensulfothion	000115-90-2	rm16561	100.4 $\pm$ 0.5 μ g/mL
heptenophos	023560-59-0	RM15898	100.3 $\pm$ 0.5 μ g/mL
isazophos	042509-80-8	RM16809	100.3 $\pm$ 0.5 μ g/mL
isocarbofos	024353-61-5	RM15033	100.4 $\pm$ 0.5 μ g/mL

Matrix: acetonitrile

Storage Conditions: Store Refrigerated (2° - 8°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality
Management System. Cert # 56 100 18560026

Page: 1 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937

Certificate of Analysis

Product Number: PSM-100-B

Lot Number: 0006565862

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality
Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937