

Certificate of Analysis

Product Name: GC Pesticide Standard #5

Product Number: PSM-100-E

Lot Issue Date: 22-Jun-2020

Lot Number: 0006543183

Expiration Date: 31-Jul-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system, and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
acrinathrin	101007-06-1	RM13927	100.5 $\pm$ 0.5 μ g/mL
bifenazate	149877-41-8	RM16270	100.3 $\pm$ 0.5 μ g/mL
bifenthrin	082657-04-3	RM15401	99.9 $\pm$ 0.5 μ g/mL
carbophenothion-methyl	000953-17-3	RM14629	100.2 $\pm$ 0.5 μ g/mL
chlorthion	000500-28-7	RM12331	100.4 $\pm$ 0.5 μ g/mL
cyfluthrin	068359-37-5	RM09294	100.4 $\pm$ 0.5 μ g/mL
cypermethrin	052315-07-8	RM07225	99.6 $\pm$ 0.5 μ g/mL
deltamethrin	052918-63-5	RM09138	100.1 $\pm$ 0.5 μ g/mL
dipropetryn	004147-51-7	RM14337	100.1 $\pm$ 0.5 μ g/mL
lambda-cyhalothrin	091465-08-6	RM16050	100.5 $\pm$ 0.5 μ g/mL
methoprotetryne	000841-06-5	RM14689	100.2 $\pm$ 0.5 μ g/mL
permethrins (mixed isomers,total)	052645-53-1	RM10021	100.1 $\pm$ 0.5 μ g/mL
prometryn	007287-19-6	RM16543	100.1 $\pm$ 0.5 μ g/mL
propazine	000139-40-2	RM14562	100.1 $\pm$ 0.5 μ g/mL
tefluthrin	079538-32-2	RM13070	100.1 $\pm$ 0.5 μ g/mL

Matrix: acetonitrile

Storage Conditions: Store Refrigerated (2° - 8°C).

Certificate of Analysis

Product Number: PSM-100-E

Lot Number: 0006543183

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert No.
AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/



ISO 17025 Cert
No. AT-1937