

Certificate of Analysis

Product Name: LC Pesticide Standard #6

Product Number: PSM-100-R

Lot Issue Date: 19-Oct-2020

Lot Number: 0006564996

Expiration Date: 30-Nov-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
acibenzolar-s-methyl	135158-54-2	RM14025	100.2 $\pm$ 0.5 μ g/mL
chlorbufam	001967-16-4	RM12387	100.3 $\pm$ 0.5 μ g/mL
ethirimol	023947-60-6	RM15486	100.3 $\pm$ 0.5 μ g/mL
formetanate hydrochloride	023422-53-9	RM02988	100.0 $\pm$ 0.5 μ g/mL
nuarimol	063284-71-9	RM14892	99.9 $\pm$ 0.5 μ g/mL
spinosad	168316-95-8	RM15747	100.1 $\pm$ 0.5 μ g/mL
spiroxamine	118134-30-8	RM14628	99.9 $\pm$ 0.5 μ g/mL
tebufenozide	112410-23-8	RM16941	100.3 $\pm$ 0.5 μ g/mL
terbuthylazine	005915-41-3	RM03389	100.0 $\pm$ 0.5 μ g/mL
thiabendazole	000148-79-8	RM09257	99.7 $\pm$ 0.5 μ g/mL
thiofanox	039196-18-4	RM16872	99.9 $\pm$ 0.5 μ g/mL
triadimenol	055219-65-3	RM03453	100.3 $\pm$ 0.5 μ g/mL
tricyclazole	041814-78-2	RM16318	100.0 $\pm$ 0.5 μ g/mL

Matrix: acetonitrile

Storage Conditions: Store Refrigerated (2° - 8°C).

Certificate of Analysis

Product Number: PSM-100-R

Lot Number: 0006564996

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality
Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937