

Certificate of Analysis

Product Name: LC Pesticide Standard #11

Product Number: PSM-100-W

Lot Issue Date: 03-Aug-2020

Lot Number: 0006550761

Expiration Date: 31-Aug-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system, and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
benomyl	017804-35-2	RM15557-01	100.0 $\pm$ 0.5 μ g/mL
clofentezine	074115-24-5	RM14474	99.9 $\pm$ 0.5 μ g/mL
clothianidin	210880-92-5	RM10326	99.8 $\pm$ 0.5 μ g/mL
fenazaquin	120928-09-8	RM14279	100.1 $\pm$ 0.5 μ g/mL
metalaxyl	057837-19-1	RM09240	100.0 $\pm$ 0.5 μ g/mL
methomyl	016752-77-5	RM16198	100.0 $\pm$ 0.5 μ g/mL
nitenpyram	150824-47-8	RM14686	100.0 $\pm$ 0.5 μ g/mL
oxadixyl	077732-09-3	RM15431	100.1 $\pm$ 0.5 μ g/mL
oxamyl	023135-22-0	RM16485	100.0 $\pm$ 0.5 μ g/mL
paclobutrazol	076738-62-0	RM05497	100.2 $\pm$ 0.5 μ g/mL
phenmedipham	013684-63-4	RM10612	99.9 $\pm$ 0.5 μ g/mL
prochloraz	067747-09-5	RM07146	100.2 $\pm$ 0.5 μ g/mL
propoxur	000114-26-1	RM16246	99.9 $\pm$ 0.5 μ g/mL
pymetrozine	123312-89-0	RM12151	100.0 $\pm$ 0.5 μ g/mL
sulfentrazone	122836-35-5	RM13708	99.9 $\pm$ 0.5 μ g/mL
thiophanate methyl	023564-05-8	RM16714	99.9 $\pm$ 0.5 μ g/mL

Matrix: acetonitrile

Certificate of Analysis

Product Number: PSM-100-W

Lot Number: 0006550761

Storage Conditions: Store Refrigerated (2° - 8°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert No.
AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/



ISO 17025 Cert
No. AT-1937