

Certificate of Analysis

Product Name: FDA Pesticide Standard GC no. 1

Product Number: PSM-101-A

Lot Issue Date: 22-Dec-2020

Lot Number: 0006579416

Expiration Date: 31-Jan-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
4,4'-DDT	000050-29-3	RM00618	100.1 $\pm$ 0.5 μ g/mL
alpha-BHC	000319-84-6	RM14886	100.0 $\pm$ 0.5 μ g/mL
beta-BHC	000319-85-7	RM17110	100.1 $\pm$ 0.5 μ g/mL
biphenyl	000092-52-4	NT01492	100.3 $\pm$ 0.5 μ g/mL
chlorpropham	000101-21-3	RM02736	100.2 $\pm$ 0.5 μ g/mL
cis-1,2,3,6-tetrahydrophthalimide	001469-48-3	RM04075	100.1 $\pm$ 0.5 μ g/mL
delta-BHC	000319-86-8	RM09156	99.9 $\pm$ 0.5 μ g/mL
dichloran	000099-30-9	RM00430	100.2 $\pm$ 0.5 μ g/mL
fenitrothion	000122-14-5	RM17644	100.1 $\pm$ 0.5 μ g/mL
metalddehyde	009002-91-9	RM09714	100.3 $\pm$ 0.5 μ g/mL
metolachlor	051218-45-2	RM14834	100.2 $\pm$ 0.5 μ g/mL
pentachloroaniline	000527-20-8	RM15381	99.9 $\pm$ 0.5 μ g/mL
pentachloroanisole	001825-21-4	RM02457	100.2 $\pm$ 0.5 μ g/mL
pentachlorobenzene	000608-93-5	RM03776	99.9 $\pm$ 0.5 μ g/mL
quintozone	000082-68-8	RM12207	100.2 $\pm$ 0.5 μ g/mL
phosalone	002310-17-0	RM16457	100.2 $\pm$ 0.5 μ g/mL
procymidone	032809-16-8	RM17154	100.0 $\pm$ 0.5 μ g/mL

Certificate of Analysis

Product Number: PSM-101-A

Lot Number: 0006579416

tecnazene	000117-18-0	RM09213	100.0 ± 0.5 µg/mL
trifluralin	001582-09-8	RM09256	100.2 ± 0.5 µg/mL
vinclozolin	050471-44-8	RM03801	100.2 ± 0.5 µg/mL

Matrix: acetone

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937