

Certificate of Analysis

Product Name: FDA Pesticide Standard GC no. 3

Product Number: PSM-101-C

Lot Issue Date: 29-Jun-2020

Lot Number: 0006544763

Expiration Date: 31-Jul-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system, and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
2,3,5,6-tetrachloroaniline	003481-20-7	NT00289	100.1 $\pm$ 0.5 μ g/mL
4,4'-DDE	000072-55-9	RM02892	100.2 $\pm$ 0.5 μ g/mL
butralin	033629-47-9	RM02706	100.2 $\pm$ 0.5 μ g/mL
captan	000133-06-2	RM02848	100.2 $\pm$ 0.5 μ g/mL
chlorfenapyr	122453-73-0	RM15241	100.1 $\pm$ 0.5 μ g/mL
cis-chlordane	005103-71-9	RM07243	100.1 $\pm$ 0.5 μ g/mL
cis-permethrin	061949-76-6	RM04820	100.3 $\pm$ 0.5 μ g/mL
cypermethrin	052315-07-8	RM07722	100.1 $\pm$ 0.5 μ g/mL
deltamethrin	052918-63-5	RM09138	100.1 $\pm$ 0.5 μ g/mL
dieldrin	000060-57-1	RM16576	100.3 $\pm$ 0.5 μ g/mL
folpet	000133-07-3	RM09175	100.3 $\pm$ 0.5 μ g/mL
lindane	000058-89-9	RM02694	100.3 $\pm$ 0.5 μ g/mL
methoxychlor	000072-43-5	RM14186	100.1 $\pm$ 0.5 μ g/mL
methyl-pentachlorophenylsulfide	001825-19-0	RM09738	100.3 $\pm$ 0.5 μ g/mL
MGK-264 (mixed, total)	000113-48-4	RM03143	100.1 $\pm$ 0.5 μ g/mL
napropamide	015299-99-7	RM15400	100.2 $\pm$ 0.5 μ g/mL
nicotine	000054-11-5	RM12837	100.2 $\pm$ 0.5 μ g/mL

Certificate of Analysis

Product Number: PSM-101-C**Lot Number:** 0006544763

o,p'-methoxychlor	030667-99-3	RM03223	100.1 ± 0.5 µg/mL
pentachlorobenzonitrile	020925-85-3	RM09889	100.1 ± 0.5 µg/mL
triallate	002303-17-5	RM03385	100.3 ± 0.5 µg/mL

Matrix: acetone**Storage Conditions:** Store at Room Temperature (15° to 30°C).**Traceability:**

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:
Monica Bourgeois
QMS RepresentativeISO 17034 Cert No.
AR-1936RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality
Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/ISO 17025 Cert
No. AT-1937