

Certificate of Analysis

Product Name: FDA Pesticide Standard LC no. 3

Product Number: PSM-101-H

Lot Issue Date: 20-Oct-2020

Lot Number: 0006565003

Expiration Date: 28-Feb-2021

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
aldicarb	000116-06-3	RM16389-01	100.0 ± 0.5 µg/mL
aldicarb sulfone	001646-88-4	RM15172	100.1 ± 0.5 µg/mL
carbetamide	016118-49-3	RM07240	100.3 ± 0.5 µg/mL
clethodim	099129-21-2	RM17286	99.9 ± 0.5 µg/mL
clofentezine	074115-24-5	RM17287	100.3 ± 0.5 µg/mL
desmedipham	013684-56-5	RM12621	100.2 ± 0.5 µg/mL
fenobucarb	003766-81-2	RM16978	100.2 ± 0.5 µg/mL
fenoxycarb	072490-01-8	RM14810	100.3 ± 0.5 µg/mL
fenuron	000101-42-8	RM02972	100.3 ± 0.5 µg/mL
forchlorfenuron	068157-60-8	RM14176	100.1 ± 0.5 µg/mL
metconazole	125116-23-6	RM12156	100.0 ± 0.5 µg/mL
pencycuron	066063-05-6	RM17317	100.3 ± 0.5 µg/mL
phenmedipham	013684-63-4	RM10604	100.1 ± 0.5 µg/mL
propoxur	000114-26-1	RM16246	100.0 ± 0.5 µg/mL
pymetrozine	123312-89-0	RM12151	100.1 ± 0.5 µg/mL
siduron (mix of isomers)	001982-49-6	RM03313	100.0 ± 0.5 µg/mL
tebuthiuron	034014-18-1	RM15610	99.9 ± 0.5 µg/mL

Certificate of Analysis

Product Number: PSM-101-H

Lot Number: 0006565003

temephos

003383-96-8

RM10837

100.2 ± 0.5 µg/mL

Matrix: acetonitrile

Storage Conditions: Store Frozen (-25° to -10°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937