

# Certificate of Analysis

**Product Name:** CA Cannabis Pesticide Mix 5

**Product Number:** PST-CBS-CA5

**Lot Issue Date:** 16-Nov-2020

**Lot Number:** 0006569788

**Expiration Date:** 31-Dec-2021

**Description:**

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

| Analyte                | CAS#        | Analyte Lot | Concentration ± Uncertainty |
|------------------------|-------------|-------------|-----------------------------|
| iprodione              | 036734-19-7 | RM16880     | 100.4 ± 0.5 µg/mL           |
| kinoprene              | 042588-37-4 | RM01521     | 100.2 ± 0.5 µg/mL           |
| methoprene             | 040596-69-8 | RM16142     | 100.4 ± 0.5 µg/mL           |
| MGK-264 (mixed, total) | 000113-48-4 | RM03143     | 100.1 ± 0.5 µg/mL           |
| novaluron              | 116714-46-6 | RM17139     | 100.4 ± 0.5 µg/mL           |
| phenothrin (tech)      | 026002-80-2 | RM17105     | 100.4 ± 0.5 µg/mL           |
| pirimicarb             | 023103-98-2 | RM11116     | 100.3 ± 0.5 µg/mL           |
| pyraclostrobin         | 175013-18-0 | RM16209     | 100.4 ± 0.5 µg/mL           |
| resmethrin             | 010453-86-8 | RM11175     | 100.3 ± 0.5 µg/mL           |
| spirodiclofen          | 148477-71-8 | RM04943     | 100.1 ± 0.5 µg/mL           |
| tebufenozide           | 112410-23-8 | RM16941     | 100.2 ± 0.5 µg/mL           |
| teflubenzuron          | 083121-18-0 | RM15412     | 100.1 ± 0.5 µg/mL           |
| tetrachlorvinphos      | 022248-79-9 | RM05693     | 100.4 ± 0.5 µg/mL           |
| tetramethrin           | 007696-12-0 | RM16228     | 100.4 ± 0.5 µg/mL           |
| thiophanate methyl     | 023564-05-8 | RM16714     | 100.4 ± 0.5 µg/mL           |

**Matrix:** acetonitrile

**Storage Conditions:** Store Frozen (-25° to -10°C).

# Certificate of Analysis

**Product Number:** PST-CBS-CA5

**Lot Number:** 0006569788

**Traceability:**

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

**Homogeneity:**

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

**Intended Use:**

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

**Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

**Hazards:**

Refer to the Safety Data Sheet on [www.agilent.com](http://www.agilent.com) for information regarding this RM.

**Expiration of Certification:**

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

**Maintenance of Certification:**

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

**Sample lot approver:**

  
Monica Bourgeois  
QMS Representative



ISO 17034 Cert  
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality  
Management System. Cert # 56 100 18560026

Page: 2 of 2

[www.agilent.com/quality/](http://www.agilent.com/quality/)  
CSD-QA-015.1



ISO 17025 Cert  
No. AT-1937