

Product Name: Octachlorostyrene

Product Number: RCB-045

Lot Issue Date: 14-May-2020

Lot Number: 0006538786

Expiration Date: 30-Apr-2023

Description:

This Reference Material (RM) was prepared in accordance with ISO 17034 and under Agilent's ISO 9001 registered quality system. The purity is reported below.

Analyte	CAS#	Purity (%)
octachlorostyrene	29082-74-4	98 ± 0.5 % (w/w)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

Balances used are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034, with a manufacturer's tolerance of ± 0.3 mg.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative