

Certificate of Analysis

Product Name: Pesticides Standard

Product Number: SCA-200-1

Lot Issue Date: 27-Aug-2020

Lot Number: 0006563434

Expiration Date: 30-Sep-2021

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
abamectin	071751-41-2	RM16771	100.1 ± 0.5 µg/mL
acephate	030560-19-1	RM16087	100.3 ± 0.5 µg/mL
acequinocyl	057960-19-7	RM11787	100.2 ± 0.5 µg/mL
acetamiprid	135410-20-7	RM15438	100.0 ± 0.5 µg/mL
aldicarb	000116-06-3	rm16389-01	100.4 ± 0.5 µg/mL
azoxystrobin	131860-33-8	RM15985	100.1 ± 0.5 µg/mL
cyfluthrin	068359-37-5	RM10927	100.2 ± 0.5 µg/mL
bifenazate	149877-41-8	RM16939	100.1 ± 0.5 µg/mL
bifenthrin	082657-04-3	RM15401	100.2 ± 0.5 µg/mL
boscalid	188425-85-6	RM09348	100.1 ± 0.5 µg/mL
carbaryl	000063-25-2	RM15786	100.2 ± 0.5 µg/mL
carbofuran	001563-66-2	RM02695	100.2 ± 0.5 µg/mL
chlorfenapyr	122453-73-0	RM15241	100.2 ± 0.5 µg/mL
chlorpyrifos	002921-88-2	RM16560	100.1 ± 0.5 µg/mL
clofentezine	074115-24-5	RM14474	100.1 ± 0.5 µg/mL
cypermethrin	052315-07-8	RM07722	100.3 ± 0.5 µg/mL
daminozide	001596-84-5	RM12081	100.2 ± 0.5 µg/mL

Certificate of Analysis

Product Number: SCA-200-1

Lot Number: 0006563434

diazinon	000333-41-5	RM16577	100.1 ± 0.5 µg/mL
chlorantraniliprole	500008-45-7	RM14473	100.1 ± 0.5 µg/mL
dichlorvos	000062-73-7	RM16942	100.1 ± 0.5 µg/mL

Matrix: acetonitrile

Storage Conditions: Store Frozen (-25° to -10°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937