

Certificate of Analysis

Product Name: Oregon Pesticides Standard 3

Product Number: SOR-120-1

Lot Issue Date: 18-Nov-2020

Lot Number: 0006573639

Expiration Date: 31-Dec-2021

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
oxamyl	023135-22-0	RM16485	100.2 $\pm$ 0.5 μ g/mL
paclobutrazol	076738-62-0	RM16797	100.0 $\pm$ 0.5 μ g/mL
permethrins (mixed isomers,total)	052645-53-1	RM10197	99.7 $\pm$ 0.5 μ g/mL
phosmet	000732-11-6	RM03255	100.3 $\pm$ 0.5 μ g/mL
piperonyl butoxide	000051-03-6	RM09150	100.5 $\pm$ 0.5 μ g/mL
propiconazole	060207-90-1	RM17177	100.2 $\pm$ 0.5 μ g/mL
propoxur	000114-26-1	RM16246	100.0 $\pm$ 0.5 μ g/mL
pyrethrins	008003-34-7	RM16467	100.1 $\pm$ 0.5 μ g/mL
pyridaben	096489-71-3	RM05236	100.2 $\pm$ 0.5 μ g/mL
spinosad	168316-95-8	RM15747	100.3 $\pm$ 0.5 μ g/mL
spiromesifen	283594-90-1	RM15237	100.3 $\pm$ 0.5 μ g/mL
spirotetramat	203313-25-1	RM17144	100.1 $\pm$ 0.5 μ g/mL
spiroxamine	118134-30-8	RM14628	100.0 $\pm$ 0.5 μ g/mL
tebuconazole	107534-96-3	RM16409	100.5 $\pm$ 0.5 μ g/mL
thiacloprid	111988-49-9	RM14388	100.2 $\pm$ 0.5 μ g/mL
thiamethoxam	153719-23-4	RM14789	100.1 $\pm$ 0.5 μ g/mL
trifloxystrobin	141517-21-7	RM16711	100.1 $\pm$ 0.5 μ g/mL

Certificate of Analysis

Product Number: SOR-120-1**Lot Number:** 0006573639

prallethrin

023031-36-9

RM16411

100.0 ± 0.5 µg/mL

Matrix: acetonitrile**Storage Conditions:** Store Frozen (-25° to -10°C).**Traceability:**

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative