

Certificate of Analysis

Product Name: Carbamate Pesticides Standard

Product Number: PPM-530-1

Lot Issue Date: 14-May-2021

Lot Number: 0006608176

Expiration Date: 30-Jun-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration $\pm$ Uncertainty
aldicarb	000116-06-3	RM16389-01	100.4 $\pm$ 0.5 $\mu\text{g/mL}$
aldicarb sulfone	001646-88-4	RM15172	100.3 $\pm$ 0.5 $\mu\text{g/mL}$
aldicarb sulfoxide	001646-87-3	RM18005	100.2 $\pm$ 0.5 $\mu\text{g/mL}$
carbaryl	000063-25-2	RM17769	100.1 $\pm$ 0.5 $\mu\text{g/mL}$
carbofuran	001563-66-2	RM17145	100.3 $\pm$ 0.5 $\mu\text{g/mL}$
3-hydroxycarbofuran	016655-82-6	RM18060	100.0 $\pm$ 0.5 $\mu\text{g/mL}$
methiocarb	002032-65-7	RM17512	100.3 $\pm$ 0.5 $\mu\text{g/mL}$
methomyl	016752-77-5	RM17393	100.2 $\pm$ 0.5 $\mu\text{g/mL}$
oxamyl	023135-22-0	RM17278	100.1 $\pm$ 0.5 $\mu\text{g/mL}$
propoxur	000114-26-1	RM16246	100.5 $\pm$ 0.5 $\mu\text{g/mL}$

Matrix: methanol (methyl alcohol)

Storage Conditions: Store Frozen (-25° to -10°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the LRQA registered ISO 9001:2015 Quality Management System. Cert # 10303760

Page: 1 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937

Certificate of Analysis

Product Number: PPM-530-1

Lot Number: 0006608176

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the LRQA registered ISO 9001:2015 Quality Management System. Cert # 10303760

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937