

Certificate of Analysis

Product Name: FDA Pesticide Standard GC/LC no. 5

Product Number: PSM-101-P

Lot Issue Date: 24-Jun-2021

Lot Number: 0006616723

Expiration Date: 31-Jul-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
chlorpyrifos-methyl	005598-13-0	RM17420	100.1 ± 0.5 µg/mL
diazinon	000333-41-5	RM18298	100.2 ± 0.5 µg/mL
ethofumesate	026225-79-6	RM00842	100.0 ± 0.5 µg/mL
ethoprop	013194-48-4	RM18225	100.2 ± 0.5 µg/mL
etoxazol	153233-91-1	RM17090	100.4 ± 0.5 µg/mL
fluopyram	658066-35-4	RM14795	100.3 ± 0.5 µg/mL
fluquinconazole	136426-54-5	RM15323	100.2 ± 0.5 µg/mL
fluridone	059756-60-4	RM13547	100.1 ± 0.5 µg/mL
flusilazole	085509-19-9	RM13159	100.2 ± 0.5 µg/mL
fluxapyroxad	907204-31-3	RM11819	100.2 ± 0.5 µg/mL
hexaconazole	079983-71-4	RM16570	100.3 ± 0.5 µg/mL
indoxacarb	173584-44-6	RM17218	100.3 ± 0.5 µg/mL
methiocarb	002032-65-7	RM17512	100.2 ± 0.5 µg/mL
metrafenone	220899-03-6	RM17191	100.2 ± 0.5 µg/mL
prometryn	007287-19-6	RM17267	100.1 ± 0.5 µg/mL
pyriproxyfen	095737-68-1	RM17394	100.1 ± 0.5 µg/mL
quinoxypfen	124495-18-7	RM14287	100.0 ± 0.5 µg/mL

Certificate of Analysis

Product Number: PSM-101-P

Lot Number: 0006616723

triadimefon	043121-43-3	RM14288	100.2 ± 0.5 µg/mL
triflumizole	068694-11-1	RM17727	100.3 ± 0.5 µg/mL

Matrix: acetonitrile

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937