

Certificate of Analysis

ICP-MS Stock Tuning Solution (100 mL)

Agilent Part Number: 5188-6564
Lot Number: 56-137CRY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.1 µg/mL	3110*	Ce ₂ (CO ₃) ₃ ·x(H ₂ O)	99.99+	Tl	7440-28-0	10.0 µg/mL	10.1 µg/mL	3158*	TlNO ₃	99.99+
Co	7440-48-4	10.0 µg/mL	10.2 µg/mL	3113*	Co	99.99+	Y	7440-65-5	10.0 µg/mL	10.1 µg/mL	3167a*	Y(NO ₃) ₃ ·6H ₂ O	99.99+
Li	7439-93-2	10.0 µg/mL	10.1 µg/mL	3129a*	Li ₂ CO ₃	99.99+							

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

 2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 14-197AB, 14-198AB, 56-136CR, 12-240AB.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.06	Cs	<0.02	Ho	0.05	Nd	<0.04	Ru	<0.05	Te	<0.03
Al	1	Cu	<0.06	In	<0.01	Ni	<0.4	Sb	<0.06	Th	<0.01
As	<0.8	Dy	0.1	Ir	0.02	P	<100	Sc	<0.2	Ti	<0.3
Au	<0.1	Er	<0.02	K	<100	Pb	0.02	Se	<1	Tm	<0.01
B	1	Eu	<0.02	La	0.1	Pd	<0.05	Si	<100	U	<0.02
Ba	<0.1	Fe	2	Lu	<0.01	Pr	0.2	Sm	<0.03	V	<0.08
Be	<0.03	Ga	<0.7	Mg	0.3	Pt	<0.02	Sn	<0.07	W	<0.03
Bi	<0.03	Gd	<0.4	Mn	<0.2	Rb	<0.01	Sr	<0.2	Yb	<0.01
Ca	7	Ge	<0.4	Mo	<0.06	Re	<0.01	Ta	<0.06	Zn	0.1
Cd	<0.04	Hf	<0.01	Na	0.1	Rh	<0.05	Tb	<0.6	Zr	<0.01
Cr	0.5	Hg	<0.01	Nb	<0.07						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

Date of release: July 15, 2021

Date of expiration: January 31, 2023



 QC Coordinator