

# Certificate of Analysis

## ICP-MS Stock Tuning Solution (100 mL)

**Agilent Part Number: 5188-6564**
**Lot Number: 57-019CRY2**

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	9.90 µg/mL	3110*	Ce <sub>2</sub> (CO <sub>3</sub> ) <sub>3</sub> ·x(H <sub>2</sub> O)	99.99+	Tl	7440-28-0	10.0 µg/mL	9.96 µg/mL	3158*	TlNO <sub>3</sub>	99.99+
Co	7440-48-4	10.0 µg/mL	9.99 µg/mL	3113*	Co	99.99+	Y	7440-65-5	10.0 µg/mL	9.99 µg/mL	3167a*	Y(NO <sub>3</sub> ) <sub>3</sub> ·6H <sub>2</sub> O	99.99+
Li	7439-93-2	10.0 µg/mL	9.94 µg/mL	3129a*	Li <sub>2</sub> CO <sub>3</sub>	99.99+							

\* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

**Purity grades:**

Starting Materials: Shown above

Matrix:

 2% HNO<sub>3</sub>: HNO<sub>3</sub> (CAS No. 7697-37-2) high purity grade

**Traceability:**

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 2-62YJ, 57-020CR, 51-238CR, 54-164CR.

**Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:**


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	0.02	Cs	<0.02	Ho	0.04	Nd	0.02	Ru	<0.1	Te	<0.2
Al	1	Cu	0.6	In	<0.01	Ni	0.4	Sb	<0.02	Th	<0.01
As	0.6	Dy	0.1	Ir	<0.02	P	<100	Sc	<0.1	Ti	0.6
Au	<0.02	Er	<0.01	K	<50	Pb	<0.04	Se	<0.2	Tm	<0.01
B	<0.5	Eu	<0.01	La	0.07	Pd	0.1	Si	<100	U	<0.03
Ba	<0.08	Fe	1	Lu	<0.01	Pr	0.3	Sm	<0.03	V	0.06
Be	<0.03	Ga	0.5	Mg	1	Pt	<0.02	Sn	<0.01	W	<0.01
Bi	<0.02	Gd	0.2	Mn	<0.2	Rb	<0.4	Sr	0.06	Yb	<0.02
Ca	<10	Ge	<0.2	Mo	<0.06	Re	<0.02	Ta	<0.01	Zn	<2
Cd	<0.01	Hf	<0.01	Na	<2	Rh	<0.01	Tb	<0.1	Zr	0.07
Cr	<0.1	Hg	<0.1	Nb	<0.03						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

**Date of release:** August 15, 2021

**Date of expiration:** February 28, 2023


  
 QC Coordinator