



Certificate of Analysis

ICP-MS Internal Std Mix

Agilent Part Number: 5188-6525

Lot Number: 57-017CRY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Bi	7440-69-9	100 µg/mL	98.8 µg/mL	3106*	Bi	99.99+	Lu	7439-94-3	100 µg/mL	99.3 µg/mL	3130a*	Lu ₂ O ₃	99.99+
Ge	7440-56-4	100 µg/mL	99.0 µg/mL	3120a*	GeO ₂	99.99+	Rh	7440-16-6	100 µg/mL	98.9 µg/mL	3144*	Rh(NO ₃) ₃ ·H ₂ O	99.99+
In	7440-74-6	100 µg/mL	98.9 µg/mL	3124a*	In	99.99+	Sc	7440-20-2	100 µg/mL	100 µg/mL	3148a*	Sc ₂ O ₃	99.99+
Li*	7439-93-2	100 µg/mL	98.8 µg/mL	3129a*	⁶ Li ₂ CO ₃	99.99+	Tb	7440-27-9	100 µg/mL	98.8 µg/mL	3157a*	Tb ₄ O ₇	99.99+

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

10% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 55-089CR, 56-139CR, 55-058CR.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	0.2	Co	0.4	Hf	0.6	Nb	0.4	Ru	<0.1	Ti	<0.01
Al	9	Cr	1	Hg	<0.5	Nd	0.07	Sb	<0.1	Tl	0.5
As	8	Cs	<0.05	Ho	0.04	Ni	0.9	Se	<10	Tm	<0.09
Au	<0.1	Cu	1	Ir	6	P	<200	Si	<200	U	<0.06
B	8	Dy	0.09	K	<100	Pb	0.9	Sm	<0.02	V	0.1
Ba	4	Er	0.06	La	0.2	Pd	<0.2	Sn	0.7	W	<0.8
Be	<0.05	Eu	0.05	Mg	6	Pr	0.02	Sr	2	Y	1
Ca	100	Fe	10	Mn	0.9	Pt	<0.03	Ta	0.02	Yb	0.6
Cd	0.2	Ga	0.2	Mo	<0.1	Rb	<0.7	Te	0.5	Zn	30
Ce	0.1	Gd	<0.04	Na	<6	Re	<0.01	Th	0.4	Zr	3

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

Date of release: September 15, 2021

Date of expiration: March 31, 2023


QC Coordinator