

# Certificate of Analysis

## ICP-MS Tuning Solution 10 ppm 100mL

**Agilent Part Number: 5190-0465**
**Lot Number: 57-037CRY2**

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	9.94 µg/mL	3110*	Ce <sub>2</sub> (CO <sub>3</sub> ) <sub>3</sub> ·x(H <sub>2</sub> O)	99.99+	Mg	7439-95-4	10.0 µg/mL	10.0 µg/mL	3131a*	Mg	99.99+
Co	7440-48-4	10.0 µg/mL	9.94 µg/mL	3113*	Co	99.99+	Tl	7440-28-0	10.0 µg/mL	9.92 µg/mL	3158*	TlNO <sub>3</sub>	99.99+
Li	7439-93-2	10.0 µg/mL	9.98 µg/mL	3129a*	Li <sub>2</sub> CO <sub>3</sub>	99.99+	Y	7440-65-5	10.0 µg/mL	9.90 µg/mL	3167a*	Y(NO <sub>3</sub> ) <sub>3</sub> ·6H <sub>2</sub> O	99.99+

\* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

**Purity grades:**

Starting Materials: Shown above

Matrix:

 2% HNO<sub>3</sub>: HNO<sub>3</sub> (CAS No. 7697-37-2) high purity grade

**Traceability:**

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 56-103CR, 3-117MJ, 1-201YJ, 3-150MKB.

**Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:**


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	0.6	Cs	<0.01	Ho	0.04	Nd	0.03	Ru	<0.05	Te	<0.2
Al	1	Cu	<0.1	In	<0.02	Ni	<0.01	Sb	<0.02	Th	<0.01
As	0.4	Dy	0.1	Ir	<0.08	P	<100	Sc	<0.2	Ti	<0.6
Au	<0.05	Er	<0.02	K	0.7	Pb	<0.02	Se	<0.6	Tm	<0.01
B	<0.3	Eu	<0.01	La	0.08	Pd	<0.2	Si	<100	U	<0.01
Ba	0.09	Fe	0.6	Lu	<0.01	Pr	0.3	Sm	<0.02	V	<0.05
Be	<0.05	Ga	<0.1	Mn	0.5	Pt	<0.04	Sn	<0.08	W	<0.2
Bi	<0.01	Gd	<0.1	Mo	<0.1	Rb	0.04	Sr	<0.05	Yb	<0.01
Ca	6	Ge	<0.2	Na	1	Re	<0.02	Ta	<0.01	Zn	<0.9
Cd	<0.01	Hf	<0.03	Nb	<0.01	Rh	<0.02	Tb	0.6	Zr	<0.1
Cr	<0.2	Hg	<0.04								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

**Date of release:** October 31, 2021

**Date of expiration:** April 30, 2023


  
 QC Coordinator