

Pesticides Matrix Spiking Solution

Product Number: CLP-200N**Page:** 1 of 1**Lot Number:** CG-3554**Lot Issue Date:** Oct-2010**Expiration Date:** Nov-2012

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
gamma-BHC	000058-89-9	NT01899	2006 ± 10 µg/mL
heptachlor	000076-44-8	224-124B	2006 ± 10 µg/mL
aldrin	000309-00-2	214-135B	2006 ± 10 µg/mL
dieldrin	000060-57-1	ER030105-03	5015 ± 25 µg/mL
endrin	000072-20-8	256-27A	5015 ± 25 µg/mL
4,4'-DDT	000050-29-3	02230KY	5015 ± 25 µg/mL

Matrix: methanol (methyl alcohol)**Storage:** Store at Room Temperature (18-25° C)

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality