

**ICP-MS Tuning Solution  
Low Concentration**
**Catalog Number: CICM-699**
**Lot Number: R00907**
**Lot Issue Date: 08/28/2013**
**Expiration Date: 09/30/2015**

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
lithium	1.0	±	0.0	µg/L	ICP / ICP-MS	3129a
yttrium	1.0	±	0.0	µg/L	ICP / ICP-MS	3167a
cerium	1.0	±	0.0	µg/L	ICP / ICP-MS	3110
thallium	1.0	±	0.0	µg/L	ICP / ICP-MS	3158
cobalt	1.0	±	0.0	µg/L	ICP / ICP-MS	3113
magnesium	1.0	±	0.0	µg/L	ICP / ICP-MS	3131a

**Matrix:** 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NC SL Z-540-1 and ISO 9001.



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851.01

ISO 9001:2000  
Registered  
TUV USA, Inc.  
Cert. No. 06-1004



William J. Leary  
Quality Assurance Manager