

Custom Standard

Product Number: CUS-5530

Page: 1 of 1

Lot Number: CH-1967

Lot Issue Date: 15-Jun-2011

Expiration Date: 31-Jul-2013

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA Scientific's ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
3-methylphenol	000108-39-4	03517AG	2002 ± 10 µg/mL
2,4-dichlorophenol	000120-83-2	05729EZ	2002 ± 10 µg/mL
2,4,6-trichlorophenol	000088-06-2	07509ME	2002 ± 10 µg/mL
4-chloro-3-methylphenol	000059-50-7	061657	2002 ± 10 µg/mL
2-nitrophenol	000088-75-5	N960045	2000 ± 10 µg/mL
4-nitrophenol	000100-02-7	HE5518HE	2008 ± 10 µg/mL
2-methyl-4,6-dinitrophenol	000534-52-1	168-60A	2004 ± 10 µg/mL
pentachlorophenol	000087-86-5	NT00751	2006 ± 10 µg/mL
phenol	000108-95-2	1373574	2004 ± 10 µg/mL
2-methylphenol	000095-48-7	09416AA	2006 ± 10 µg/mL
4-methylphenol	000106-44-5	09410PI	2008 ± 10 µg/mL

Matrix: methanol/methylene chloride (1:1)

Storage: Store at Room Temperature (18-25° C)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com



William J. Leary
Quality Assurance Manager

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality