

Certificate of Analysis

ISO Guide 34

Pesticides Individual Standard

Product Number: CLP-216D-1**Page:** 1 of 1**Lot Number:** 0006468594**Lot Issue Date:** 24-May-2019**Expiration Date:** 30-Jun-2022

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alpha-BHC	000319-84-6	RM14073	20.1 ± 0.1 ng/mL
gamma-BHC	000058-89-9	RM02694	20.0 ± 0.1 ng/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM01256	40.0 ± 0.2 ng/mL
dieldrin	000060-57-1	RM13088	40.1 ± 0.2 ng/mL
endosulfan I	000959-98-8	RM00834	20.0 ± 0.1 ng/mL
endrin	000072-20-8	RM00557	40.0 ± 0.2 ng/mL
4,4'-DDD	000072-54-8	RM00615	40.1 ± 0.2 ng/mL
4,4'-DDT	000050-29-3	NT01175	40.1 ± 0.2 ng/mL
heptachlor	000076-44-8	RM03111	20.1 ± 0.1 ng/mL
methoxychlor	000072-43-5	RM12528	200.0 ± 1.0 ng/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	RM13844	20.0 ± 0.1 ng/mL

Matrix: hexane**Storage:** Store at Room Temperature (15° to 30°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS RepresentativeISO Guide 34 Cert No.
AR-1936Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management SystemISO17025 Cert No.
AT-1937