

Certificate of Analysis

ISO 17034

Pesticides Performance Evaluation Standard

Product Number: CLP-252-1**Page:** 1 of 1**Lot Number:** 0006482220**Lot Issue Date:** 09-Aug-2019**Expiration Date:** 30-Sep-2022

This ISO 17034 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alpha-BHC	000319-84-6	RM10865	10.0 ± 0.1 ng/mL
beta-BHC	000319-85-7	RM06986	10.0 ± 0.1 ng/mL
gamma-BHC	000058-89-9	RM02694	10.0 ± 0.1 ng/mL
4,4'-DDT	000050-29-3	RM06888	100.3 ± 0.5 ng/mL
decachlorobiphenyl (BZ # 209)	002051-24-3	RM01256	20.1 ± 0.1 ng/mL
endrin	000072-20-8	RM00557	50.2 ± 0.3 ng/mL
methoxychlor	000072-43-5	RM10269	250.3 ± 1.3 ng/mL
2,4,5,6-tetrachloro-m-xylene	000877-09-8	RM09245	20.1 ± 0.1 ng/mL

Matrix: hexane**Storage:** Store at Room Temperature (15° to 30°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS Representative